

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074100.D
 Acq On : 24 Aug 2022 18:44
 Operator : JC\MD
 Sample : N4280-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20220817

Quant Time: Aug 25 07:55:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	566882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	856893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	718931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	292337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	291483	52.791	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.580%			
35) Dibromofluoromethane	7.951	113	257244	50.763	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.520%			
50) Toluene-d8	10.380	98	966639	49.252	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.500%			
62) 4-Bromofluorobenzene	12.674	95	291126	44.695	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.400%			
Target Compounds						Qvalue
3) Chloromethane	2.269	50	1094	0.219	ug/l #	72
9) 1,1,2-Trichlorotrifluo...	4.146	101	12399	2.579	ug/l	97
12) 1,1-Dichloroethene	4.110	96	12066	2.697	ug/l	88
27) cis-1,2-Dichloroethene	7.251	96	8018	1.318	ug/l	86
30) Chloroform	7.751	83	3963	0.405	ug/l	96
32) 1,1,1-Trichloroethane	7.945	97	1839	0.212	ug/l #	45
38) Carbon Tetrachloride	8.139	117	9294	1.155	ug/l	92
44) Trichloroethene	9.151	130	811526	126.477	ug/l	99
55) 1,1,2-Trichloroethane	10.833	97	3102	0.523	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074100.D
Acq On : 24 Aug 2022 18:44
Operator : JC\MD
Sample : N4280-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D2-20220817

Quant Time: Aug 25 07:55:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

