

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074101.D
 Acq On : 24 Aug 2022 19:08
 Operator : JC\MD
 Sample : N4280-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT14911-20220817

Quant Time: Aug 25 07:56:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	562965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	855619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	719087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	307806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	292824	53.403	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.800%
35) Dibromofluoromethane	7.951	113	259209	51.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.460%
50) Toluene-d8	10.380	98	951189	48.537	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.080%
62) 4-Bromofluorobenzene	12.674	95	294503	45.281	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.560%
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	5.540	73	3620	0.226	ug/l	92
44) Trichloroethene	9.157	130	3357	0.524	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074101.D
Acq On : 24 Aug 2022 19:08
Operator : JC\MD
Sample : N4280-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT149I1-20220817

Quant Time: Aug 25 07:56:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

