

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074103.D
 Acq On : 24 Aug 2022 19:58
 Operator : JC\MD
 Sample : N4280-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D3-20220818

Quant Time: Aug 25 07:56:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	571999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	874119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	730587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	306542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	295102	52.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.940%
35) Dibromofluoromethane	7.957	113	264187	51.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.220%
50) Toluene-d8	10.380	98	980478	48.973	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.940%
62) 4-Bromofluorobenzene	12.674	95	295060	44.406	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.820%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.145	101	5793	1.194	ug/l #	78
44) Trichloroethene	9.151	130	970803	148.319	ug/l	99
64) Tetrachloroethene	10.916	164	1796	0.311	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074103.D
Acq On : 24 Aug 2022 19:58
Operator : JC\MD
Sample : N4280-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D3-20220818

Quant Time: Aug 25 07:56:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

