

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082423\
 Data File : VN079024.D
 Acq On : 24 Aug 2023 11:47
 Operator : JC\MD
 Sample : VN0824MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBL01

Quant Time: Aug 25 01:10:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	182907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	346682	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	306450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92156	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	143201	50.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.620%
35) Dibromofluoromethane	8.171	113	120500	51.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.280%
50) Toluene-d8	10.571	98	413059	47.669	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.340%
62) 4-Bromofluorobenzene	12.853	95	120710	42.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082423\
Data File : VN079024.D
Acq On : 24 Aug 2023 11:47
Operator : JC\MD
Sample : VN0824MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBL01

Quant Time: Aug 25 01:10:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

