

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082423\
 Data File : VN079032.D
 Acq On : 24 Aug 2023 15:29
 Operator : JC\MD
 Sample : 04075-01 20X
 Misc : 6.52g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NEVINS-SB05-Z113-114

Quant Time: Aug 25 02:54:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	249619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	449475	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	418253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	174972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	175499	45.182	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.360%
35) Dibromofluoromethane	8.171	113	148841	49.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	10.570	98	531256	47.288	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.580%
62) 4-Bromofluorobenzene	12.853	95	192661	52.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.940%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	11.970	91	32452	2.004	ug/l	90
68) m/p-Xylenes	12.076	106	7812	1.291	ug/l	90
84) 1,2,4-Trimethylbenzene	13.488	105	24233	2.117	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082423\
Data File : VN079032.D
Acq On : 24 Aug 2023 15:29
Operator : JC\MD
Sample : 04075-01 20X
Misc : 6.52g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NEVINS-SB05-Z113-114

Quant Time: Aug 25 02:54:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

