

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082520\
 Data File : VN063117.D
 Acq On : 25 Aug 2020 17:27
 Operator : JC/MD
 Sample : L3787-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1498

Quant Time: Aug 26 05:40:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	171207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	300844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	283670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125127	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	109282	53.86	ug/l	0.00
Spiked Amount			Recovery	=	107.72%	
35) Dibromofluoromethane	7.55	113	90167	46.00	ug/l	0.00
Spiked Amount			Recovery	=	92.00%	
50) Toluene-d8	10.06	98	363093	48.22	ug/l	0.00
Spiked Amount			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	12.38	95	131414	49.46	ug/l	0.00
Spiked Amount			Recovery	=	98.92%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.04	50	1672	0.712	ug/l	95
11) Tert butyl alcohol	4.74	59	16151	56.194	ug/l	99
16) Acetone	3.79	43	37153	55.667	ug/l	98
17) Carbon Disulfide	4.02	76	3097	0.814	ug/l #	84
18) Methyl Acetate	4.29	43	78398	38.879	ug/l	100
20) Methylene Chloride	4.51	84	1065	Below Cal		90
25) 2-Butanone	6.80	43	11612	10.779	ug/l	92
31) Cyclohexane	7.62	56	1300	0.435	ug/l #	1
36) 1,1-Dichloropropene	7.63	75	12783	4.822	ug/l #	49
37) Ethyl Acetate	6.80	43	11612	4.674	ug/l #	68
38) Carbon Tetrachloride	7.63	117	15541	5.806	ug/l #	17
40) Benzene	8.00	78	1863	0.222	ug/l	98
43) Isopropyl Acetate	8.27	43	94081	22.860	ug/l #	53
51) 4-Methyl-2-Pentanone	9.95	43	1523	0.585	ug/l #	80
52) Toluene	10.12	92	1426	0.281	ug/l	87
59) 2-Hexanone	10.73	43	660	0.364	ug/l #	52
70) Styrene	11.94	104	461	1.989	ug/l #	36
71) Bromoform	12.38	173	926	0.525	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	12.49	83	63	Below Cal	#	26
76) 1,2,3-Trichloropropane	12.38	75	62724	25.789	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	62724	70.829	ug/l #	15
89) n-Butylbenzene	13.59	91	794	2.881	ug/l #	62
93) 1,2,4-Trichlorobenzene	14.88	180	442	4.730	ug/l #	86
95) Naphthalene	15.11	128	1760	5.955	ug/l #	71
96) 1,2,3-Trichlorobenzene	15.28	180	347	4.116	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082520\
Data File : VN063117.D
Acq On : 25 Aug 2020 17:27
Operator : JC/MD
Sample : L3787-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1498

Quant Time: Aug 26 05:40:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. 0.00 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27

Instrument :

MSVOA_N

ClientSampleId :


1498

Tot Ion:168 Resp: 171207

Ion Ratio Lower Upper

168 100

99 53.0 42.0 63.0



Abundance Ion 167.90 (167.60 to 168.60): VN062996.D (-1803) (-)

Ion 98.90 (98.60 to 99.60): VN062996.D (-1803) (-)

7.63

80000

60000



40000

20000

0

Time-->

7.50 7.60 7.70



#3

Chloromethane

Concen: 0.712 ug/l

RT: 2.04 min Scan# 83

Delta R.T. 0.01 min

Lab File: VN063117.D


Acq: 25 Aug 2020 17:27

Tgt Ion: 50 Resp: 1672

Ion Ratio Lower Upper

50 100

52 35.4 26.0 39.0



Abundance Ion 49.90 (49.60 to 50.60): VN062996.D (-70) (-)

Ion 51.90 (51.60 to 52.60): VN062996.D (-70) (-)

2.04

1000

800

600


400


200

0

Time-->

2.00 2.05









#11

Tert butyl alcohol
 Concen: 56.194 ug/l
 RT: 4.74 min Scan# 923
 Delta R.T. -0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Instrument :
 MSVOA_N
 ClientSampleId :
 1498




Tot Ion: 59 Resp: 16151
 Ion Ratio Lower Upper
 59 100
 57 10.0 7.8 11.6




#16

Acetone
 Concen: 55.667 ug/l
 RT: 3.79 min Scan# 626
 Delta R.T. 0.01 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Tgt Ion: 43 Resp: 37153
 Ion Ratio Lower Upper
 43 100
 58 33.7 27.8 41.6





#17

Carbon Disulfide

Concen: 0.814 ug/l

RT: 4.02 min Scan# 698

Delta R.T. 0.01 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27

Instrument :

MSVOA_N

ClientSampled :


1498

Tot Ion: 76 Resp: 3097

Ion Ratio Lower Upper




76 100

78 14.4 6.9 10.3#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0



#18

Methyl Acetate

Concen: 38.879 ug/l

RT: 4.29 min Scan# 782

Delta R.T. 0.01 min

Lab File: VN063117.D


Acq: 25 Aug 2020 17:27

Tgt Ion: 43 Resp: 78398

Ion Ratio Lower Upper



43 100


74 23.4 18.6 28.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#20
Methvlene Chloride
Concen: Below Cal
RT: 4.51 min Scan# 851
Delta R.T. 0.01 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Instrument :
MSVOA_N
ClientSampleId :
1498

Tot Ion: 84 Resp: 1065
Ion Ratio Lower Upper
84 100
49 138.2 98.7 148.1
51 42.9 30.8 46.2
86 58.7 51.2 76.8




Abundance Ion 83.90 (83.60 to 84.60): VNC
Ion 48.90 (48.60 to 49.60): VNC
Ion 50.90 (50.60 to 51.60): VNC
Ion 85.90 (85.60 to 86.60): VNC






#25
2-Butanone
Concen: 10.779 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. 0.01 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 43 Resp: 11612
Ion Ratio Lower Upper
43 100
72 23.3 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VNC
Ion 72.00 (71.70 to 72.70): VNC









#31
Cyclohexane
Concen: 0.435 ug/l
RT: 7.62 min Scan# 1817
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Instrument :
MSVOA_N
ClientSampled :
1498

Tot Ion: 56 Resp: 1300
Ion Ratio Lower Upper
56 100
69 190.0 26.6 40.0#
84 134.7 66.4 99.6#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 53.857 ug/l
RT: 7.99 min Scan# 1934
Delta R.T. 0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 65 Resp: 109282
Ion Ratio Lower Upper
65 100
67 54.2 0.0 112.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2109
 Delta R.T. 0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Instrument :
 MSVOA_N
 ClientSampleId :
 1498

Tot Ion:114 Resp: 300844
 Ion Ratio Lower Upper
 114 100
 63 21.4 0.0 43.2
 88 16.3 0.0 30.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 45.997 ug/l
 RT: 7.55 min Scan# 1797
 Delta R.T. 0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Tgt Ion:113 Resp: 90167
 Ion Ratio Lower Upper
 113 100
 111 103.1 81.7 122.5
 192 21.6 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 4.822 ug/l
 RT: 7.63 min Scan# 1821
 Delta R.T. -0.13 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Instrument :
 MSVOA_N
 ClientSampleId :
 1498




Tot Ion	75	Resp	12783
Ion Ratio	Lower	Upper	
75	100		
110	7.3	17.5	52.6#
77	0.0	24.2	36.4#



Abundance Ion 74.90 (74.60 to 75.60): VN063117.D (-1767) (-)


Ion 109.90 (109.60 to 110.60): VN063117.D (-1767) (-)

Ion 76.90 (76.60 to 77.60): VN063117.D (-1767) (-)



#37
 Ethyl Acetate
 Concen: 4.674 ug/l
 RT: 6.80 min Scan# 1563
 Delta R.T. -0.08 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27



Tgt Ion	43	Resp	11612
Ion Ratio	Lower	Upper	
43	100		
61	0.0	11.8	17.8#
70	0.0	8.2	12.2#




Abundance Ion 43.00 (42.70 to 43.70): VN063117.D (-1496) (-)

Ion 60.90 (60.60 to 61.60): VN063117.D (-1496) (-)

Ion 70.00 (69.70 to 70.70): VN063117.D (-1496) (-)









#38
Carbon Tetrachloride
Concen: 5.806 ug/l
RT: 7.63 min Scan# 1821
Delta R.T. -0.11 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Instrument :
MSVOA_N
ClientSampled :
1498

Tot Ion:117 Resp: 15541
Ion Ratio Lower Upper
117 100
119 5.9 76.2 114.4#
121 0.0 24.2 36.2#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#40
Benzene
Concen: 0.222 ug/l
RT: 8.00 min Scan# 1938
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 78 Resp: 1863
Ion Ratio Lower Upper
78 100
77 22.7 18.8 28.2



Abundance Ion 78.00 (77.70 to 78.70): V
Ion 77.00 (76.70 to 77.70): V









#43
Isopropyl Acetate
Concen: 22.860 ug/l
RT: 8.27 min Scan# 2021
Delta R.T. 0.14 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27




Instrument :
MSVOA_N
ClientSampled :
1498


Tot Ion: 43 Resp: 94081
Ion Ratio Lower Upper
43 100
61 0.1 22.6 33.8#
87 0.0 10.5 15.7#



#50
Toluene-d8
Concen: 48.222 ug/l
RT: 10.06 min Scan# 2578
Delta R.T. 0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 98 Resp: 363093
Ion Ratio Lower Upper
98 100
100 64.1 51.2 76.8









#51
4-Methyl-2-Pentanone
Concen: 0.585 ug/l
RT: 9.95 min Scan# 2544
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27




Instrument :
MSVOA_N
ClientSampleId :
1498


Tot Ion: 43 Resp: 1523
Ion Ratio Lower Upper
43 100
58 27.5 32.2 48.2#



#52
Toluene
Concen: 0.281 ug/l
RT: 10.12 min Scan# 2597
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 92 Resp: 1426
Ion Ratio Lower Upper
92 100
91 191.7 139.2 208.8









#59
2-Hexanone
Concen: 0.364 ug/l
RT: 10.73 min Scan# 2786
Delta R.T. 0.01 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27




Instrument :
MSVOA_N
ClientSampled :
1498


Tot Ion: 43 Resp: 660
Ion Ratio Lower Upper
43 100
58 91.5 28.1 84.3#



#62
4-Bromofluorobenzene
Concen: 49.458 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion: 95 Resp: 131414
Ion Ratio Lower Upper
95 100
174 85.8 0.0 170.8
176 81.3 0.0 164.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2988

Delta R.T. -0.00 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27

Instrument :

MSVOA_N

ClientSampleId :

1498


Tot Ion:117 Resp: 283670

Ion Ratio Lower Upper

117 100

82 55.0 44.5 66.7




119 31.8 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): V



#70

Styrene

Concen: 1.989 ug/l

RT: 11.94 min Scan# 3163

Delta R.T. 0.00 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27


Tgt Ion:104 Resp: 461

Ion Ratio Lower Upper

104 100

78 0.0 39.0 58.6#



103 13.7 44.3 66.5#




Abundance Ion 104.00 (103.70 to 104.70): V

Ion 78.00 (77.70 to 78.70): VN0

Ion 103.00 (102.70 to 103.70): V









#71
Bromoform
Concen: 0.525 ug/l
RT: 12.38 min Scan# 3299
Delta R.T. 0.28 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Instrument :
MSVOA_N
ClientSampleId :
1498

Tot Ion:173 Resp: 926
Ion Ratio Lower Upper
173 100
175 736.7 24.3 72.8#
254 0.0 0.0 0.0




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3591
Delta R.T. -0.00 min
Lab File: VN063117.D
Acq: 25 Aug 2020 17:27

Tgt Ion:152 Resp: 125127
Ion Ratio Lower Upper
152 100
115 56.9 28.7 86.3
150 157.6 0.0 348.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.49 min Scan# 3334

Delta R.T. 0.01 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27

Instrument :

MSVOA_N

ClientSampleId :

1498


Tot Ion: 83 Resp: 63

Ion Ratio Lower Upper

83 100

131 0.0 4.9 14.7#

85 0.0 31.9 95.5#




Abundance Ion 82.90 (82.60 to 83.60): VN063117.D (-3237) (-)


Ion 130.80 (130.50 to 131.50): VN063117.D (-3237) (-)

Ion 84.90 (84.60 to 85.60): VN063117.D (-3237) (-)

Time-->



Time-->



#76

1,2,3-Trichloropropane

Concen: 25.789 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN063117.D


Acq: 25 Aug 2020 17:27

Tgt Ion: 75 Resp: 62724

Ion Ratio Lower Upper

75 100


77 1.3 0.0 0.0#




Abundance Ion 74.90 (74.60 to 75.60): VN063117.D (-3252) (-)

Ion 77.00 (76.70 to 77.70): VN063117.D (-3252) (-)

Time-->



Time-->



#81

trans-1,4-Dichloro-2-butene

Concen: 70.829 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27

Instrument :

MSVOA_N

ClientSampleId :

1498





Tot Ion: 75 Resp: 62724

Ion Ratio Lower Upper

75 100

53 0.0 69.2 103.8#

89 0.0 34.2 51.4#



#89

n-Butylbenzene

Concen: 2.881 ug/l

RT: 13.59 min Scan# 3676

Delta R.T. 0.00 min

Lab File: VN063117.D

Acq: 25 Aug 2020 17:27




Tgt Ion: 91 Resp: 794


Ion Ratio Lower Upper

91 100

92 29.3 25.9 77.6

134 0.0 12.4 37.0#









#93
 1,2,4-Trichlorobenzene
 Concen: 4.730 ug/l
 RT: 14.88 min Scan# 4077
 Delta R.T. -0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27




Instrument :
 MSVOA_N
 ClientSampled :
 1498


Tot Ion:180 Resp: 442
 Ion Ratio Lower Upper
 180 100
 182 92.3 47.0 141.2
 145 0.0 14.1 42.4#



#95
 Naphthalene
 Concen: 5.955 ug/l
 RT: 15.11 min Scan# 4147
 Delta R.T. 0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Tgt Ion:128 Resp: 1760
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.6 16.0#
 129 1.7 8.9 13.3#






#96
 1,2,3-Trichlorobenzene
 Concen: 4.116 ug/l
 RT: 15.28 min Scan# 4201
 Delta R.T. -0.00 min
 Lab File: VN063117.D
 Acq: 25 Aug 2020 17:27

Instrument :
 MSVOA_N
 ClientSampleId :
 1498

Tot Ion	Ratio	Lower	Upper
180	100		
182	84.1	47.3	141.9
145	0.0	15.6	46.7



Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.90 (144.60 to 145.60): V

