

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082520\
 Data File : VN063117.D
 Acq On : 25 Aug 2020 17:27
 Operator : JC/MD
 Sample : L3787-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1498

Quant Time: Aug 26 06:45:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	171207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	300844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	283670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	109282	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
35) Dibromofluoromethane	7.55	113	90167	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
50) Toluene-d8	10.06	98	363093	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	12.38	95	131414	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
Target Compounds						
11) Tert butyl alcohol	4.74	59	16151	56.194	ug/l	99
16) Acetone	3.79	43	37153	55.667	ug/l	98
18) Methyl Acetate	4.29	43	78398	38.879	ug/l	100
25) 2-Butanone	6.80	43	11612	10.779	ug/l	92
52) Toluene	10.12	92	1426	0.281	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082520\
Data File : VN063117.D
Acq On : 25 Aug 2020 17:27
Operator : JC/MD
Sample : L3787-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1498

Quant Time: Aug 26 06:45:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

