

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
 Data File : VN068294.D
 Acq On : 25 Aug 2021 12:14
 Operator : JC/MD
 Sample : VN0825WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 26 04:00:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	507466	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	836744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	835602	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	326722	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	295954	54.621	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.240%
35) Dibromofluoromethane	8.021	113	269827	51.363	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.720%
50) Toluene-d8	10.443	98	1039487	51.930	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.860%
62) 4-Bromofluorobenzene	12.734	95	340130	51.043	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
Data File : VN068294.D
Acq On : 25 Aug 2021 12:14
Operator : JC/MD
Sample : VN0825WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 26 04:00:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

