

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
 Data File : VN068298.D
 Acq On : 25 Aug 2021 14:22
 Operator : JC/MD
 Sample : M3445-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 26 05:08:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	441108	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	736888	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	717150	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	266402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225229	47.821	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.640%
35) Dibromofluoromethane	8.024	113	204198	44.138	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	88.280%
50) Toluene-d8	10.443	98	794767	45.085	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	90.180%
62) 4-Bromofluorobenzene	12.733	95	253483	43.195	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
Data File : VN068298.D
Acq On : 25 Aug 2021 14:22
Operator : JC/MD
Sample : M3445-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 26 05:08:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

