

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
 Data File : VN068308.D
 Acq On : 25 Aug 2021 18:55
 Operator : JC/MD
 Sample : M3529-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31113

Manual Integrations
 APPROVED

MMDadoda
 8/26/2021 9:06:04 AM

Quant Time: Aug 26 05:40:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	501606	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	797062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	894022	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	556337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	234713	43.824	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	87.640%		
35) Dibromofluoromethane	8.024	113	216692	43.302	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	86.600%		
50) Toluene-d8	10.443	98	901072	47.257	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.520%		
62) 4-Bromofluorobenzene	12.734	95	398429	62.769	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	125.540%		
Target Compounds						Qvalue
11) Tert butyl alcohol	5.390	59	75137	112.497	ug/l	100
16) Acetone	4.291	43	1131926	723.865	ug/l #	86
17) Carbon Disulfide	4.521	76	6976m	0.664	ug/l	
25) 2-Butanone	7.343	43	151322	62.095	ug/l #	87
39) Methylcyclohexane	9.464	83	14531	1.975	ug/l	88
40) Benzene	8.461	78	730970	36.756	ug/l	98
51) 4-Methyl-2-Pentanone	10.333	43	81876	14.990	ug/l	88
52) Toluene	10.508	92	913592	67.387	ug/l	99
59) 2-Hexanone	11.087	43	28219	7.441	ug/l	98
64) Tetrachloroethene	10.982	164	3815	0.552	ug/l #	85
67) Ethyl Benzene	11.848	91	322217	10.672	ug/l	93
68) m/p-Xylenes	11.950	106	988660	81.853	ug/l	79
69) o-Xylene	12.283	106	509790	43.682	ug/l	84
70) Styrene	12.296	104	850799	45.546	ug/l #	88
73) Isopropylbenzene	12.583	105	54003	1.365	ug/l	99
78) n-propylbenzene	12.924	91	337517	7.952	ug/l #	69
80) 1,3,5-Trimethylbenzene	13.061	105	318145	9.939	ug/l	90
84) 1,2,4-Trimethylbenzene	13.369	105	1473223	47.147	ug/l	93
85) sec-Butylbenzene	13.503	105	28970m	0.774	ug/l	
86) p-Isopropyltoluene	13.619	119	46520	1.506	ug/l	88
87) 1,3-Dichlorobenzene	13.619	146	11631m	0.626	ug/l	
88) 1,4-Dichlorobenzene	13.696	146	15242	0.809	ug/l	73
89) n-Butylbenzene	13.943	91	40948	1.777	ug/l #	83
91) 1,2-Dichlorobenzene	13.991	146	13681	0.773	ug/l	85
93) 1,2,4-Trichlorobenzene	15.265	180	36447	6.401	ug/l	97
94) Hexachlorobutadiene	15.375	225	4112	0.858	ug/l	94
95) Naphthalene	15.493	128	23989824m	1028.388	ug/l	
96) 1,2,3-Trichlorobenzene	15.702	180	46972	7.860	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082521\
Data File : VN068308.D
Acq On : 25 Aug 2021 18:55
Operator : JC/MD
Sample : M3529-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
31113

Manual Integrations
APPROVED

MMDadoda
8/26/2021 9:06:04 AM

Quant Time: Aug 26 05:40:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion: 168 Resp: 501606
Ion Ratio Lower Upper
168 100
99 50.2 62.9 94.3#

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#11
Tert butyl alcohol
Concen: 112.497 ug/l
RT: 5.390 min Scan# 1268
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55


Tgt Ion: 59 Resp: 75137
Ion Ratio Lower Upper
59 100
57 10.9 8.7 13.1





#16
Acetone
Concen: 723.865 ug/l
RT: 4.291 min Scan# 858
Delta R.T. 0.003 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion: 43 Resp: 1131926
Ion Ratio Lower Upper
43 100
58 33.7 21.2 31.8#

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#17
Carbon Disulfide
Concen: 0.664 ug/l m
RT: 4.521 min Scan# 944
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion: 76 Resp: 6976
Ion Ratio Lower Upper
76 100
78 9.6 7.7 11.5









#25
2-Butanone
Concen: 62.095 ug/l
RT: 7.343 min Scan# 1996
Delta R.T. 0.003 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion: 43 Resp: 151322
Ion Ratio Lower Upper
43 100
72 30.8 19.7 29.5#

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#33
1,2-Dichloroethane-d4
Concen: 43.824 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion: 65 Resp: 234713
Ion Ratio Lower Upper
65 100
67 51.2 0.0 98.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.971 min Scan# 2603
Delta R.T. 0.003 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:114 Resp: 797062
Ion Ratio Lower Upper
114 100
63 17.3 0.0 49.0
88 14.3 0.0 36.6

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#35
Dibromofluoromethane
Concen: 43.302 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.003 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:113 Resp: 216692
Ion Ratio Lower Upper
113 100
111 101.7 83.8 125.8
192 20.9 13.4 20.0#









#39
Methylcyclohexane
Concen: 1.975 ug/l
RT: 9.464 min Scan# 2787
Delta R.T. 0.002 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion: 83 Resp: 14531
Ion Ratio Lower Upper
83 100
55 68.5 66.6 99.8
98 47.1 35.3 52.9

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#40
Benzene
Concen: 36.756 ug/l
RT: 8.461 min Scan# 2413
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55


Tgt Ion: 78 Resp: 730970
Ion Ratio Lower Upper
78 100
77 24.5 18.9 28.3





#50
Toluene-d8
Concen: 47.257 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion: 98 Resp: 901072
Ion Ratio Lower Upper
98 100
100 64.5 51.7 77.5

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#51
4-Methyl-2-Pentanone
Concen: 14.990 ug/l
RT: 10.333 min Scan# 3111
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion: 43 Resp: 81876
Ion Ratio Lower Upper
43 100
58 42.7 28.6 43.0









#52
Toluene
Concen: 67.387 ug/l
RT: 10.508 min Scan# 3176
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion: 92 Resp: 913592
Ion Ratio Lower Upper
92 100
91 175.8 139.6 209.4

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#59
2-Hexanone
Concen: 7.441 ug/l
RT: 11.087 min Scan# 3392
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55


Tgt Ion: 43 Resp: 28219
Ion Ratio Lower Upper
43 100
58 50.8 24.7 74.1





#62
4-Bromofluorobenzene
Concen: 62.769 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion: 95 Resp: 398429
Ion Ratio Lower Upper
95 100
174 105.9 0.0 122.0
176 100.8 0.0 117.6

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion: 117 Resp: 894022
Ion Ratio Lower Upper
117 100
82 49.1 53.1 79.7#
119 33.3 25.5 38.3









#64
Tetrachloroethene
Concen: 0.552 ug/l
RT: 10.982 min Scan# 3353
Delta R.T. 0.002 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:	164	Resp:	3815
Ion	Ratio	Lower	Upper
164	100		
166	125.0	102.1	153.1
129	94.7	86.9	130.3
131	70.3	83.6	125.4#

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#67
Ethyl Benzene
Concen: 10.672 ug/l
RT: 11.848 min Scan# 3676
Delta R.T. 0.002 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:	91	Resp:	322217
Ion	Ratio	Lower	Upper
91	100		
106	33.0	23.4	35.0









#68
m/p-Xylenes
Concen: 81.853 ug/l
RT: 11.950 min Scan# 3714
Delta R.T. -0.006 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:106 Resp: 988660
Ion Ratio Lower Upper
106 100
91 190.7 180.6 270.8

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#69
o-Xylene
Concen: 43.682 ug/l
RT: 12.283 min Scan# 3838
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:106 Resp: 509790
Ion Ratio Lower Upper
106 100
91 210.2 118.3 354.9









#70
Styrene
Concen: 45.546 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:104 Resp: 850799
Ion Ratio Lower Upper
104 100
78 44.1 48.2 72.2#
103 56.4 44.5 66.7

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55


Tgt Ion:152 Resp: 556337
Ion Ratio Lower Upper
152 100
115 60.1 36.1 108.5
150 155.9 0.0 351.8





#73
Isopropylbenzene
Concen: 1.365 ug/l
RT: 12.583 min Scan# 3950
Delta R.T. 0.002 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion:105 Resp: 54003
Ion Ratio Lower Upper
105 100
120 24.5 12.5 37.5

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#78
n-propylbenzene
Concen: 7.952 ug/l
RT: 12.924 min Scan# 4077
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion: 91 Resp: 337517
Ion Ratio Lower Upper
91 100
120 6.5 10.4 31.4#









#80
1,3,5-Trimethylbenzene
Concen: 9.939 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55

Tgt Ion:105 Resp: 318145
Ion Ratio Lower Upper
105 100
120 51.1 22.2 66.6

Instrument :
MSVOA_N
ClientSampled :
31113




Manual Integrations
APPROVED


MMDadoda
8/26/2021 9:06:04 AM



#84
1,2,4-Trimethylbenzene
Concen: 47.147 ug/l
RT: 13.369 min Scan# 4243
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55


Tgt Ion:105 Resp: 1473223
Ion Ratio Lower Upper
105 100
120 45.5 20.6 61.8





#85
 sec-Butylbenzene
 Concen: 0.774 ug/l m
 RT: 13.503 min Scan# 4293
 Delta R.T. 0.000 min
 Lab File: VN068308.D
 Acq: 25 Aug 2021 18:55




Instrument :
 MSVOA_N
 ClientSampled :
 31113




Tgt Ion:105 Resp: 28970
 Ion Ratio Lower Upper
 105 100
 134 0.0 9.0 27.1#

Manual Integrations
 APPROVED



MMDadoda
 8/26/2021 9:06:04 AM




#86
 p-Isopropyltoluene
 Concen: 1.506 ug/l
 RT: 13.619 min Scan# 4336
 Delta R.T. -0.000 min
 Lab File: VN068308.D
 Acq: 25 Aug 2021 18:55




Tgt Ion:119 Resp: 46520
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.4 37.4
 91 40.3 14.8 44.3





#87
1,3-Dichlorobenzene
Concen: 0.626 ug/l m
RT: 13.619 min Scan# 4336
Delta R.T. -0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion:146 Resp: 11631
Ion Ratio Lower Upper
146 100
111 89.7 24.3 72.8#
148 106.9 32.0 96.0#

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#88
1,4-Dichlorobenzene
Concen: 0.809 ug/l
RT: 13.696 min Scan# 4365
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion:146 Resp: 15242
Ion Ratio Lower Upper
146 100
111 68.1 23.4 70.2
148 81.6 31.8 95.3





#89

n-Butylbenzene

Concen: 1.777 ug/l

RT: 13.943 min Scan# 4457

Delta R.T. -0.003 min

Lab File: VN068308.D





Acq: 25 Aug 2021 18:55

Tgt Ion:	91	Resp:	40948
Ion	Ratio	Lower	Upper
91	100		
92	53.8	25.7	77.1
134	0.0	11.2	33.6#

Instrument :
MSVOA_N
ClientSampled :
31113

Manual Integrations
APPROVED

MMDadoda
8/26/2021 9:06:04 AM



#91

1,2-Dichlorobenzene

Concen: 0.773 ug/l




RT: 13.991 min Scan# 4475


Delta R.T. 0.000 min

Lab File: VN068308.D

Acq: 25 Aug 2021 18:55


Tgt Ion:	146	Resp:	13681
Ion	Ratio	Lower	Upper
146	100		
111	30.8	24.9	74.6
148	67.5	31.6	95.0





#93
1,2,4-Trichlorobenzene
Concen: 6.401 ug/l
RT: 15.265 min Scan# 4950
Delta R.T. 0.000 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion:180 Resp: 36447
Ion Ratio Lower Upper
180 100
182 95.6 47.5 142.5
145 39.9 17.2 51.6

Manual Integrations
APPROVED



MMDadoda
8/26/2021 9:06:04 AM




#94
Hexachlorobutadiene
Concen: 0.858 ug/l
RT: 15.375 min Scan# 4991
Delta R.T. 0.002 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Tgt Ion:225 Resp: 4112
Ion Ratio Lower Upper
225 100
223 68.3 31.4 94.0
227 66.3 31.2 93.6





#95
Naphthalene
Concen: 1028.388 ug/l m
RT: 15.493 min Scan# 5035
Delta R.T. -0.014 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55




Instrument :
MSVOA_N
ClientSampled :
31113




Tgt Ion:128 Resp:23989824
Ion Ratio Lower Upper
128 100
127 0.0 10.6 15.8#
129 0.0 8.6 12.8#

Manual Integrations
APPROVED

MMDadoda
8/26/2021 9:06:04 AM



#96
1,2,3-Trichlorobenzene
Concen: 7.860 ug/l
RT: 15.702 min Scan# 5113
Delta R.T. -0.001 min
Lab File: VN068308.D
Acq: 25 Aug 2021 18:55



Tgt Ion:180 Resp: 46972
Ion Ratio Lower Upper
180 100
182 95.3 47.8 143.3
145 56.5 18.2 54.6#

