

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
 Data File : VN074108.D
 Acq On : 25 Aug 2022 12:28
 Operator : JC\MD
 Sample : VN0825WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0825WBL01

Quant Time: Aug 26 01:57:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	588216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	886983	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	729176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	297106	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	303137	52.911	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.820%
35) Dibromofluoromethane	7.951	113	266052	50.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.440%
50) Toluene-d8	10.381	98	983666	48.420	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.840%
62) 4-Bromofluorobenzene	12.674	95	291409	43.221	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
Data File : VN074108.D
Acq On : 25 Aug 2022 12:28
Operator : JC\MD
Sample : VN0825WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0825WBL01

Quant Time: Aug 26 01:57:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

