

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
 Data File : VN074112.D
 Acq On : 25 Aug 2022 14:20
 Operator : JC\MD
 Sample : N4244-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-103D

Quant Time: Aug 26 02:00:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	583577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	874924	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	723545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	297875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	298338	52.487	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.980%
35) Dibromofluoromethane	7.951	113	268807	51.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.900%
50) Toluene-d8	10.380	98	971284	48.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	12.668	95	292946	44.048	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.100%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.110	96	3341	0.725	ug/l	# 89
24) 1,1-Dichloroethane	6.298	63	17063	1.895	ug/l	95
27) cis-1,2-Dichloroethene	7.245	96	7861	1.256	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
Data File : VN074112.D
Acq On : 25 Aug 2022 14:20
Operator : JC\MD
Sample : N4244-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-103D

Quant Time: Aug 26 02:00:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

