

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
 Data File : VN074119.D
 Acq On : 25 Aug 2022 17:13
 Operator : JC\MD
 Sample : N4283-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 337

Quant Time: Aug 26 02:03:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	619880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	938657	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	822249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	367417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	316125	52.359	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.720%	
35) Dibromofluoromethane	7.951	113	279984	50.437	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.880%	
50) Toluene-d8	10.380	98	1072221	49.873	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.740%	
62) 4-Bromofluorobenzene	12.674	95	343016	48.074	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	71085	24.956	ug/l	98
20) Methylene Chloride	5.022	84	34131	5.678	ug/l #	86
43) Isopropyl Acetate	8.498	43	64760	4.867	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
Data File : VN074119.D
Acq On : 25 Aug 2022 17:13
Operator : JC\MD
Sample : N4283-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
337

Quant Time: Aug 26 02:03:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

