

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
 Data File : VN074123.D
 Acq On : 25 Aug 2022 18:52
 Operator : JC\MD
 Sample : N4280-10DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D2-20220818DL

Quant Time: Aug 26 02:05:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	586312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	894789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	737599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	323547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	305755	53.541	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.080%
35) Dibromofluoromethane	7.957	113	269424	50.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	10.380	98	986674	48.144	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.280%
62) 4-Bromofluorobenzene	12.674	95	302657	44.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.000%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.151	130	341158	50.918	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082522\
Data File : VN074123.D
Acq On : 25 Aug 2022 18:52
Operator : JC\MD
Sample : N4280-10DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D2-20220818DL

Quant Time: Aug 26 02:05:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

