

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082523\
 Data File : VN079053.D
 Acq On : 25 Aug 2023 16:32
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Aug 25 23:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	243031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	446710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	414289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	152005	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	184357	48.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	8.171	113	157081	52.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.480%
50) Toluene-d8	10.571	98	535172	47.932	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.853	95	178330	48.864	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082523\
Data File : VN079053.D
Acq On : 25 Aug 2023 16:32
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Aug 25 23:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

