

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082525\
 Data File : VN087664.D
 Acq On : 25 Aug 2025 17:29
 Operator : JC\MD
 Sample : Q2909-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4065

Quant Time: Aug 26 01:40:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	174514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	396343	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	363540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	168797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	185252	51.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.620%	
35) Dibromofluoromethane	8.147	113	139292	48.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.360%	
50) Toluene-d8	10.547	98	509350	47.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.120%	
62) 4-Bromofluorobenzene	12.829	95	178923	46.974	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.940%	
Target Compounds						Qvalue
16) Acetone	4.418	43	17526	9.004	ug/l	94
52) Toluene	10.612	92	151939	18.202	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082525\
Data File : VN087664.D
Acq On : 25 Aug 2025 17:29
Operator : JC\MD
Sample : Q2909-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4065

Quant Time: Aug 26 01:40:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

