

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082525\
 Data File : VN087666.D
 Acq On : 25 Aug 2025 18:10
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 01:41:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.710	0.661	6.9	75	0.00
3 P	Chloromethane	0.730	0.650	11.0	75	0.00
4 C	Vinyl Chloride	0.846	0.748	11.6#	76	0.00
5 T	Bromomethane	0.437	0.365	16.5	75	0.00
6 T	Chloroethane	0.595	0.509	14.5	78	0.00
7 T	Trichlorofluoromethane	1.240	1.151	7.2	83	0.00
8 T	Diethyl Ether	0.540	0.460	14.8	80	0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.701	0.615	12.3	81	0.00
10 T	Methyl Iodide	0.450	0.311	30.9#	59	0.00
11 T	Tert butyl alcohol	0.178	0.157	11.8	78	0.00
12 CM	1,1-Dichloroethene	0.721	0.625	13.3#	82	0.00
13 T	Acrolein	0.219	0.209	4.6	90	0.00
14 T	Allyl chloride	1.306	1.034	20.8	77	0.01
15 T	Acrylonitrile	0.502	0.432	13.9	80	0.00
16 T	Acetone	0.558	0.411	26.3#	72	0.00
17 T	Carbon Disulfide	1.985	1.648	17.0	76	0.00
18 T	Methyl Acetate	1.168	1.042	10.8	83	0.00
19 T	Methyl tert-butyl Ether	2.704	2.392	11.5	80	0.00
20 T	Methylene Chloride	0.948	0.707	25.4#	79	0.00
21 T	trans-1,2-Dichloroethene	0.781	0.651	16.6	80	0.00
22 T	Diisopropyl ether	2.819	2.470	12.4	79	0.00
23 T	Vinyl Acetate	2.565	2.292	10.6	78	0.00
24 P	1,1-Dichloroethane	1.546	1.310	15.3	80	0.01
25 T	2-Butanone	0.734	0.619	15.7	78	0.00
26 T	2,2-Dichloropropane	1.327	1.060	20.1	74	0.00
27 T	cis-1,2-Dichloroethene	0.934	0.792	15.2	77	0.00
28 T	Bromochloromethane	0.747	0.688	7.9	90	0.00
29 T	Tetrahydrofuran	0.472	0.409	13.3	79	0.00
30 C	Chloroform	1.578	1.383	12.4#	80	0.00
31 T	Cyclohexane	1.289	1.077	16.4	80	0.00
32 T	1,1,1-Trichloroethane	1.337	1.147	14.2	81	0.00
33 S	1,2-Dichloroethane-d4	1.024	0.883	13.8	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.361	0.322	10.8	85	0.00
36 T	1,1-Dichloropropene	0.554	0.481	13.2	81	0.00
37 T	Ethyl Acetate	0.757	0.635	16.1	76	0.00
38 T	Carbon Tetrachloride	0.547	0.496	9.3	81	0.00
39 T	Methylcyclohexane	0.610	0.523	14.3	79	0.00
40 TM	Benzene	1.699	1.508	11.2	79	0.00
41 T	Methacrylonitrile	0.392	0.357	8.9	81	0.00
42 TM	1,2-Dichloroethane	0.653	0.584	10.6	80	0.00
43 T	Isopropyl Acetate	1.182	1.077	8.9	80	0.00
44 TM	Trichloroethene	0.388	0.329	15.2	77	0.00
45 C	1,2-Dichloropropane	0.448	0.388	13.4#	79	0.00
46 T	Dibromomethane	0.319	0.287	10.0	80	0.00
47 T	Bromodichloromethane	0.653	0.573	12.3	79	0.00
48 T	Methyl methacrylate	0.559	0.501	10.4	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082525\
 Data File : VN087666.D
 Acq On : 25 Aug 2025 18:10
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 01:41:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	78	0.00
50 S	Toluene-d8	1.351	1.196	11.5	87	0.00
51 T	4-Methyl-2-Pentanone	0.713	0.671	5.9	81	0.00
52 CM	Toluene	1.053	0.922	12.4#	78	0.00
53 T	t-1,3-Dichloropropene	0.676	0.610	9.8	77	0.00
54 T	cis-1,3-Dichloropropene	0.702	0.629	10.4	78	0.00
55 T	1,1,2-Trichloroethane	0.407	0.377	7.4	82	0.00
56 T	Ethyl methacrylate	0.651	0.613	5.8	77	0.00
57 T	1,3-Dichloropropane	0.721	0.656	9.0	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.352	0.267	24.1	67	0.00
59 T	2-Hexanone	0.451	0.457	-1.3	79	0.00
60 T	Dibromochloromethane	0.472	0.416	11.9	79	0.00
61 T	1,2-Dibromoethane	0.430	0.385	10.5	79	0.00
62 S	4-Bromofluorobenzene	0.481	0.433	10.0	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.347	0.280	19.3	81	0.00
65 PM	Chlorobenzene	1.244	1.066	14.3	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.413	0.360	12.8	77	0.00
67 C	Ethyl Benzene	2.154	1.912	11.2#	80	0.00
68 T	m/p-Xylenes	0.790	0.704	10.9	78	0.00
69 T	o-Xylene	0.774	0.679	12.3	77	0.00
70 T	Styrene	1.297	1.186	8.6	77	0.00
71 P	Bromoform	0.317	0.285	10.1	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	4.040	3.499	13.4	79	0.00
74 T	N-amyl acetate	1.531	1.345	12.1	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.391	1.241	10.8	83	0.00
76 T	1,2,3-Trichloropropane	1.173	1.151	1.9	98	0.00
77 T	Bromobenzene	0.953	0.811	14.9	78	0.00
78 T	n-propylbenzene	4.977	4.342	12.8	79	0.00
79 T	2-Chlorotoluene	2.995	2.619	12.6	80	0.00
80 T	1,3,5-Trimethylbenzene	3.428	3.014	12.1	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.443	0.393	11.3	79	0.00
82 T	4-Chlorotoluene	3.149	2.679	14.9	77	0.00
83 T	tert-Butylbenzene	2.856	2.508	12.2	78	0.00
84 T	1,2,4-Trimethylbenzene	3.431	3.042	11.3	79	0.00
85 T	sec-Butylbenzene	4.239	3.723	12.2	80	0.00
86 T	p-Isopropyltoluene	3.500	3.055	12.7	79	0.00
87 T	1,3-Dichlorobenzene	1.876	1.619	13.7	81	0.00
88 T	1,4-Dichlorobenzene	1.952	1.612	17.4	79	0.00
89 T	n-Butylbenzene	3.476	2.999	13.7	79	0.00
90 T	Hexachloroethane	0.667	0.579	13.2	80	0.00
91 T	1,2-Dichlorobenzene	1.780	1.521	14.6	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.330	0.273	17.3	78	0.00
93 T	1,2,4-Trichlorobenzene	1.069	0.902	15.6	80	0.00
94 T	Hexachlorobutadiene	0.381	0.318	16.5	81	0.00
95 T	Naphthalene	3.786	3.299	12.9	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082525\
Data File : VN087666.D
Acq On : 25 Aug 2025 18:10
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 26 01:41:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.045	0.864	17.3	78	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6