

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057526.D
 Acq On : 26 Aug 2019 22:04
 Operator : JC/SP
 Sample : K4493-30
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-J-(0)

Quant Time: Aug 27 08:17:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	372410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	657033	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	569247	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	142737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	307261	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	7.58	113	205389	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	10.08	98	875331	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	244426	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
Target Compounds						
16) Acetone	3.80	43	9798	3.257	ug/l #	88
18) Methyl Acetate	4.31	43	21554	2.412	ug/l #	86
20) Methylene Chloride	4.52	84	10075	2.187	ug/l	89
43) Isopropyl Acetate	8.15	43	89417	7.351	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057526.D
Acq On : 26 Aug 2019 22:04
Operator : JC/SP
Sample : K4493-30
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-J-(0)

Quant Time: Aug 27 08:17:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

