

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057531.D
 Acq On : 27 Aug 2019 00:11
 Operator : JC/SP
 Sample : K4493-36
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-K4-(0)

Quant Time: Aug 27 08:27:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	366199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	666334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	560264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	138371	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	322238	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
35) Dibromofluoromethane	7.57	113	211698	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	10.09	98	858793	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	12.40	95	236962	41.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.18%	
Target Compounds						
18) Methyl Acetate	4.31	43	21672	2.488	ug/l	96
20) Methylene Chloride	4.53	84	10539	2.331	ug/l	91
43) Isopropyl Acetate	8.16	43	93269	7.561	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057531.D
Acq On : 27 Aug 2019 00:11
Operator : JC/SP
Sample : K4493-36
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-K4-(0)

Quant Time: Aug 27 08:27:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

