

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057533.D
 Acq On : 27 Aug 2019 1:02
 Operator : JC/SP
 Sample : K4493-38
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-K2-(1.9)

Quant Time: Aug 27 08:38:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	378126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	677998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	132498	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	329074	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
35) Dibromofluoromethane	7.58	113	213307	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	10.08	98	859227	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
62) 4-Bromofluorobenzene	12.40	95	229429	39.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.20%	
Target Compounds						
16) Acetone	3.79	43	8437	2.762	ug/l #	89
18) Methyl Acetate	4.31	43	21768	2.394	ug/l	98
20) Methylene Chloride	4.52	84	10088	2.156	ug/l	96
43) Isopropyl Acetate	8.16	43	83557	6.657	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057533.D
Acq On : 27 Aug 2019 1:02
Operator : JC/SP
Sample : K4493-38
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-K2-(1.9)

Quant Time: Aug 27 08:38:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

