

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057535.D
 Acq On : 27 Aug 2019 1:53
 Operator : JC/SP
 Sample : K4493-40
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-K4-(1.9)

Quant Time: Aug 27 08:41:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	382380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	676772	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	547241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	128177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	323382	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
35) Dibromofluoromethane	7.58	113	219655	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
50) Toluene-d8	10.08	98	861063	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
62) 4-Bromofluorobenzene	12.40	95	227898	38.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.82%	
Target Compounds						
16) Acetone	3.80	43	9830	3.182	ug/l	99
20) Methylene Chloride	4.53	84	9789	2.066	ug/l #	85
43) Isopropyl Acetate	8.16	43	92617	7.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057535.D
Acq On : 27 Aug 2019 1:53
Operator : JC/SP
Sample : K4493-40
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-K4-(1.9)

Quant Time: Aug 27 08:41:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

