

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057536.D
 Acq On : 27 Aug 2019 2:18
 Operator : JC/SP
 Sample : K4493-41
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-G2-(0)

Quant Time: Aug 27 09:15:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	402425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	704334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	573402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	139147	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	337068	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
35) Dibromofluoromethane	7.57	113	229643	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	10.08	98	877872	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.40	95	233569	38.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.62%	
Target Compounds						
16) Acetone	3.80	43	21372	6.574	ug/l #	86
18) Methyl Acetate	4.31	43	21032	2.084	ug/l #	91
20) Methylene Chloride	4.53	84	9560	1.911	ug/l #	84
43) Isopropyl Acetate	8.16	43	63598	4.877	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057536.D
Acq On : 27 Aug 2019 2:18
Operator : JC/SP
Sample : K4493-41
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-G2-(0)

Quant Time: Aug 27 09:15:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

