

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082619\
 Data File : VN057535.D
 Acq On : 27 Aug 2019 1:53
 Operator : JC/SP
 Sample : K4493-40
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-K4-(1.9)

Quant Time: Aug 27 08:41:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	382380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	676772	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	547241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	128177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	323382	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
35) Dibromofluoromethane	7.58	113	219655	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
50) Toluene-d8	10.08	98	861063	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
62) 4-Bromofluorobenzene	12.40	95	227898	38.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.82%	
Target Compounds						
16) Acetone	3.80	43	9830	3.182	ug/l	99
20) Methylene Chloride	4.53	84	9789	2.066	ug/l #	85
43) Isopropyl Acetate	8.16	43	92617	7.392	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082619\
Data File : VN057535.D
Acq On : 27 Aug 2019 1:53
Operator : JC/SP
Sample : K4493-40
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-K4-(1.9)

Quant Time: Aug 27 08:41:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. -0.00 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :


T9-12-13-K4-(1.9)

Tot Ion:168 Resp: 382380

Ion Ratio Lower Upper




168 100

99 67.5 50.6 76.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#16

Acetone

Concen: 3.182 ug/l

RT: 3.80 min Scan# 629

Delta R.T. -0.00 min

Lab File: VN057535.D


Acq: 27 Aug 2019 1:53

Tgt Ion: 43 Resp: 9830

Ion Ratio Lower Upper



43 100


58 27.9 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN





#20

Methylene Chloride

Concen: 2.066 ug/l

RT: 4.53 min Scan# 858

Delta R.T. 0.01 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

T9-12-13-K4-(1.9)

Tot Ion: 84 Resp: 9789


Ion Ratio Lower Upper

84 100

49 116.3 117.0 175.4#

51 47.4 35.5 53.3

86 59.8 49.8 74.8






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 48.076 ug/l

RT: 8.01 min Scan# 1941

Delta R.T. -0.00 min

Lab File: VN057535.D


Acq: 27 Aug 2019 1:53

Tgt Ion: 65 Resp: 323382

Ion Ratio Lower Upper



65 100


67 51.1 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.57 min Scan# 2115
Delta R.T. -0.00 min
Lab File: VN057535.D
Acq: 27 Aug 2019 1:53

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-K4-(1.9)

Tot Ion:114 Resp: 676772
Ion Ratio Lower Upper
114 100
63 23.9 0.0 50.2
88 17.2 0.0 34.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 48.051 ug/l
RT: 7.58 min Scan# 1805
Delta R.T. 0.00 min
Lab File: VN057535.D
Acq: 27 Aug 2019 1:53

Tgt Ion:113 Resp: 219655
Ion Ratio Lower Upper
113 100
111 100.8 83.8 125.6
192 16.0 13.5 20.3



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 7.392 ug/l

RT: 8.16 min Scan# 1985

Delta R.T. -0.00 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

T9-12-13-K4-(1.9)


Tot Ion: 43 Resp: 92617

Ion Ratio Lower Upper

43 100

61 18.1 14.5 21.7




87 10.2 8.6 12.8



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: 49.815 ug/l

RT: 10.08 min Scan# 2585

Delta R.T. -0.00 min

Lab File: VN057535.D


Acq: 27 Aug 2019 1:53

Tgt Ion: 98 Resp: 861063

Ion Ratio Lower Upper



98 100


100 63.7 50.4 75.6



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#62

4-Bromofluorobenzene

Concen: 38.906 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

T9-12-13-K4-(1.9)


Tot Ion: 95 Resp: 227898

Ion Ratio Lower Upper

95 100

174 67.8 0.0 133.0




176 65.7 0.0 130.8



Abundance Ion 94.90 (94.60 to 95.60): VN057535.D (-3212) (-)

Ion 173.80 (173.50 to 174.50): VN057535.D (-3212) (-)

Ion 175.80 (175.50 to 176.50): VN057535.D (-3212) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53


Tgt Ion: 117 Resp: 547241

Ion Ratio Lower Upper

117 100

82 63.4 50.6 75.8



119 32.2 24.6 36.8




Abundance Ion 116.90 (116.60 to 117.60): VN057535.D (-2902) (-)

Ion 82.00 (81.70 to 82.70): VN057535.D (-2902) (-)

Ion 118.90 (118.60 to 119.60): VN057535.D (-2902) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN057535.D

Acq: 27 Aug 2019 1:53

Instrument :

MSVOA_N

ClientSampleId :

T9-12-13-K4-(1.9)


Tot Ion:152 Resp: 128177

Ion Ratio Lower Upper

152 100

115 61.5 32.0 96.0

150 158.3 0.0 346.6



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

