


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN082619\
Data File : VN057497.D
Acq On : 26 Aug 2019 8:45
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Title : SW846 8260
Last Update : Fri Aug 23 05:07:29 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.6	15190	PASS
75	95	30	60	58.3	34669	PASS
95	95	100	100	100.0	59416	PASS
96	95	5	9	7.3	4317	PASS
173	174	0.00	2	0.5	205	PASS
174	95	50	100	70.1	41629	PASS
175	174	5	9	8.3	3435	PASS
176	174	95	101	97.4	40549	PASS
177	176	5	9	7.2	2936	PASS