

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082620\
 Data File : VN063136.D
 Acq On : 26 Aug 2020 16:35
 Operator : JC/MD
 Sample : L3788-03 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1794

Quant Time: Aug 27 00:54:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	154053	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	258060	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	262407	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	113396	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	96904	53.07	ug/l	0.00
Spiked Amount			Recovery	=	106.14%	
35) Dibromofluoromethane	7.55	113	84254	50.11	ug/l	0.00
Spiked Amount			Recovery	=	100.22%	
50) Toluene-d8	10.06	98	329041	50.94	ug/l	0.00
Spiked Amount			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	12.38	95	118956	52.19	ug/l	0.00
Spiked Amount			Recovery	=	104.38%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.55	94	309	0.255	ug/l	83
11) Tert butyl alcohol	4.73	59	22489	86.958	ug/l	98
16) Acetone	3.78	43	124977	219.019	ug/l	98
20) Methylene Chloride	4.51	84	7919	3.682	ug/l	96
25) 2-Butanone	6.79	43	38547	39.767	ug/l	99
31) Cyclohexane	7.63	56	2834	1.053	ug/l #	1
36) 1,1-Dichloropropene	7.63	75	11237	4.941	ug/l #	45
37) Ethyl Acetate	6.79	43	38547	18.089	ug/l #	68
38) Carbon Tetrachloride	7.63	117	14230	6.198	ug/l #	17
39) Methylcyclohexane	9.05	83	92	1.591	ug/l #	18
48) Methyl methacrylate	9.00	41	1982	1.187	ug/l #	38
51) 4-Methyl-2-Pentanone	9.95	43	1921	0.861	ug/l #	75
59) 2-Hexanone	10.73	43	4330	2.782	ug/l	83
70) Styrene	11.94	104	115	1.941	ug/l #	26
71) Bromoform	12.38	173	872	0.534	ug/l #	1
76) 1,2,3-Trichloropropane	12.38	75	56162	25.480	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	56162	69.980	ug/l #	15
89) n-Butylbenzene	13.59	91	926	2.912	ug/l #	79
90) Hexachloroethane	13.92	117	539	0.389	ug/l #	5
93) 1,2,4-Trichlorobenzene	14.88	180	596	4.821	ug/l #	87
94) Hexachlorobutadiene	14.99	225	593	0.453	ug/l	73
95) Naphthalene	15.11	128	1757	5.987	ug/l #	69
96) 1,2,3-Trichlorobenzene	15.29	180	616	4.260	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082620\
Data File : VN063136.D
Acq On : 26 Aug 2020 16:35
Operator : JC/MD
Sample : L3788-03 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1794

Quant Time: Aug 27 00:54:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. 0.00 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampleId :


1794

Tot Ion:168 Resp: 154053

Ion Ratio Lower Upper

168 100

99 53.1 42.0 63.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



60000

40000

20000

0

Time--> 7.50 7.60 7.70 7.80



#5

Bromomethane

Concen: 0.255 ug/l

RT: 2.55 min Scan# 242

Delta R.T. 0.04 min

Lab File: VN063136.D


Acq: 26 Aug 2020 16:35

Tgt Ion: 94 Resp: 309

Ion Ratio Lower Upper

94 100

96 77.9 75.8 113.8



Abundance Ion 93.90 (93.60 to 94.60): VN

Ion 95.90 (95.60 to 96.60): VN

250

200


150


100

50

0

Time--> 2.52 2.54 2.56 2.58









#11
Tert butyl alcohol
Concen: 86.958 ug/l
RT: 4.73 min Scan# 921
Delta R.T. -0.01 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35




Instrument :
MSVOA_N
ClientSampled :
1794


Tot Ion: 59 Resp: 22489
Ion Ratio Lower Upper
59 100
57 10.3 7.8 11.6



#16
Acetone
Concen: 219.019 ug/l
RT: 3.78 min Scan# 623
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion: 43 Resp: 124977
Ion Ratio Lower Upper
43 100
58 33.6 27.8 41.6









#20
Methvlene Chloride
Concen: 3.682 ug/l
RT: 4.51 min Scan# 850
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35




Instrument :
MSVOA_N
ClientSampleId :
1794


Tot Ion:	84	Resp:	7919
Ion	Ratio	Lower	Upper
84	100		
49	117.5	98.7	148.1
51	38.3	30.8	46.2
86	59.9	51.2	76.8



#25
2-Butanone
Concen: 39.767 ug/l
RT: 6.79 min Scan# 1561
Delta R.T. 0.01 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion:	43	Resp:	38547
Ion	Ratio	Lower	Upper
43	100		
72	27.7	21.8	32.6









#31
Cyclohexane
Concen: 1.053 ug/l
RT: 7.63 min Scan# 1821
Delta R.T. 0.01 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Instrument :
MSVOA_N
ClientSampleId :
1794

Tot Ion: 56 Resp: 2834
Ion Ratio Lower Upper
56 100
69 176.1 26.6 40.0#
84 161.0 66.4 99.6#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 53.074 ug/l
RT: 7.99 min Scan# 1933
Delta R.T. -0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion: 65 Resp: 96904
Ion Ratio Lower Upper
65 100
67 54.3 0.0 112.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.55 min Scan# 2108
Delta R.T. -0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Instrument :
MSVOA_N
ClientSampleId :
1794

Tot Ion:114 Resp: 258060
Ion Ratio Lower Upper
114 100
63 20.9 0.0 43.2
88 15.8 0.0 30.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 50.106 ug/l
RT: 7.55 min Scan# 1797
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion:113 Resp: 84254
Ion Ratio Lower Upper
113 100
111 103.2 81.7 122.5
192 22.1 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.941 ug/l
 RT: 7.63 min Scan# 1821
 Delta R.T. -0.13 min
 Lab File: VN063136.D
 Acq: 26 Aug 2020 16:35

Instrument :
 MSVOA_N
 ClientSampled :
 1794

Tot Ion: 75 Resp: 11237
 Ion Ratio Lower Upper
 75 100
 110 3.3 17.5 52.6#
 77 0.0 24.2 36.4#




Abundance Ion 74.90 (74.60 to 75.60): VN0
 Ion 109.90 (109.60 to 110.60): VN0
 Ion 76.90 (76.60 to 77.60): VN0






#37
 Ethyl Acetate
 Concen: 18.089 ug/l
 RT: 6.79 min Scan# 1561
 Delta R.T. -0.09 min
 Lab File: VN063136.D
 Acq: 26 Aug 2020 16:35

Tgt Ion: 43 Resp: 38547
 Ion Ratio Lower Upper
 43 100
 61 0.0 11.8 17.8#
 70 0.0 8.2 12.2#



Abundance Ion 43.00 (42.70 to 43.70): VN0
 Ion 60.90 (60.60 to 61.60): VN0
 Ion 70.00 (69.70 to 70.70): VN0





#38

Carbon Tetrachloride

Concen: 6.198 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. -0.11 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampled :

1794


Tot Ion:117 Resp: 14230

Ion Ratio Lower Upper

117 100

119 6.5 76.2 114.4#




121 0.0 24.2 36.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#39

Methylcyclohexane

Concen: 1.591 ug/l

RT: 9.05 min Scan# 2264

Delta R.T. 0.01 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35


Tgt Ion: 83 Resp: 92

Ion Ratio Lower Upper

83 100

55 0.0 63.3 94.9#



98 0.0 34.2 51.4#




Abundance Ion 83.00 (82.70 to 83.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 98.00 (97.70 to 98.70): V









#48
Methvl methacrylate
Concen: 1.187 ug/l
RT: 9.00 min Scan# 2249
Delta R.T. -0.16 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35




Instrument :
MSVOA_N
ClientSampled :
1794


Tot Ion: 41 Resp: 1982
Ion Ratio Lower Upper
41 100
69 0.0 74.2 111.4#
39 53.3 42.5 63.7



#50
Toluene-d8
Concen: 50.945 ug/l
RT: 10.06 min Scan# 2578
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion: 98 Resp: 329041
Ion Ratio Lower Upper
98 100
100 64.0 51.2 76.8





#51

4-Methyl-2-Pentanone

Concen: 0.861 ug/l

RT: 9.95 min Scan# 2544

Delta R.T. -0.00 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampled :


1794

Tot Ion: 43 Resp: 1921

Ion Ratio Lower Upper




43 100

58 24.9 32.2 48.2#



Abundance Ion 43.00 (42.70 to 43.70): VN062996.D (-2528) (-)

Ion 58.00 (57.70 to 58.70): VN062996.D (-2528) (-)



#59

2-Hexanone

Concen: 2.782 ug/l

RT: 10.73 min Scan# 2785

Delta R.T. 0.00 min

Lab File: VN063136.D


Acq: 26 Aug 2020 16:35

Tgt Ion: 43 Resp: 4330

Ion Ratio Lower Upper



43 100


58 43.7 28.1 84.3



Abundance Ion 43.00 (42.70 to 43.70): VN062996.D (-2767) (-)

Ion 58.00 (57.70 to 58.70): VN062996.D (-2767) (-)









#62
4-Bromofluorobenzene
Concen: 52.192 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. -0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Instrument :
MSVOA_N
ClientSampled :
1794

Tot Ion: 95 Resp: 118956
Ion Ratio Lower Upper
95 100
174 87.4 0.0 170.8
176 83.9 0.0 164.8




Abundance Ion 94.90 (94.60 to 95.60): VN063136.D (-3205) (-)
Ion 173.80 (173.50 to 174.50): VN063136.D (-3205) (-)
Ion 175.80 (175.50 to 176.50): VN063136.D (-3205) (-)






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion: 117 Resp: 262407
Ion Ratio Lower Upper
117 100
82 54.8 44.5 66.7
119 31.7 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): VN063136.D (-2895) (-)
Ion 82.00 (81.70 to 82.70): VN063136.D (-2895) (-)
Ion 118.90 (118.60 to 119.60): VN063136.D (-2895) (-)









#70
Stvrene
Concen: 1.941 ug/l
RT: 11.94 min Scan# 3161
Delta R.T. -0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Instrument :
MSVOA_N
ClientSampled :
1794

Tot Ion:104 Resp: 115
Ion Ratio Lower Upper
104 100
78 0.0 39.0 58.6#
103 0.0 44.3 66.5#




Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): V
Ion 103.00 (102.70 to 103.70): V






#71
Bromoform
Concen: 0.534 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.28 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion:173 Resp: 872
Ion Ratio Lower Upper
173 100
175 722.9 24.3 72.8#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. -0.00 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampleId :

1794


Tot Ion:152 Resp: 113396

Ion Ratio Lower Upper

152 100

115 56.9 28.7 86.3




150 157.5 0.0 348.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 25.480 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN063136.D


Acq: 26 Aug 2020 16:35

Tgt Ion: 75 Resp: 56162

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 69.980 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampleId :

1794


Tot Ion: 75 Resp: 56162

Ion Ratio Lower Upper

75 100

53 0.0 69.2 103.8#




89 0.0 34.2 51.4#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#89

n-Butylbenzene

Concen: 2.912 ug/l

RT: 13.59 min Scan# 3675

Delta R.T. 0.00 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35


Tgt Ion: 91 Resp: 926

Ion Ratio Lower Upper

91 100

92 56.7 25.9 77.6



134 0.0 12.4 37.0#




Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 92.00 (91.70 to 92.70): VN0

Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 0.389 ug/l

RT: 13.92 min Scan# 3779

Delta R.T. 0.07 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35

Instrument :

MSVOA_N

ClientSampled :


1794

Tot Ion:117 Resp: 539

Ion Ratio Lower Upper

117 100

201 0.0 44.1 132.4#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V

13.92



300

200

100

0

Time--> 13.88 13.90 13.92 13.94



#93

1,2,4-Trichlorobenzene

Concen: 4.821 ug/l

RT: 14.88 min Scan# 4077

Delta R.T. -0.00 min

Lab File: VN063136.D

Acq: 26 Aug 2020 16:35


Tgt Ion:180 Resp: 596

Ion Ratio Lower Upper

180 100

182 94.8 47.0 141.2

145 0.0 14.1 42.4#



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V

14.88

400


300


200

100

0

Time--> 14.86 14.88 14.90 14.92









#94
Hexachlorobutadiene
Concen: 0.453 ug/l
RT: 14.99 min Scan# 4110
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Instrument :
MSVOA_N
ClientSampled :
1794

Tot Ion:225 Resp: 593
Ion Ratio Lower Upper
225 100
223 44.4 30.9 92.5
227 39.5 31.9 95.9




Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V






#95
Naphthalene
Concen: 5.987 ug/l
RT: 15.11 min Scan# 4148
Delta R.T. 0.00 min
Lab File: VN063136.D
Acq: 26 Aug 2020 16:35

Tgt Ion:128 Resp: 1757
Ion Ratio Lower Upper
128 100
127 0.0 10.6 16.0#
129 0.0 8.9 13.3#



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V






#96
 1,2,3-Trichlorobenzene
 Concen: 4.260 ug/l
 RT: 15.29 min Scan# 4204
 Delta R.T. 0.01 min
 Lab File: VN063136.D
 Acq: 26 Aug 2020 16:35

Instrument :
 MSVOA_N
 ClientSampleId :
 1794

Tot Ion	Ratio	Lower	Upper
180	100		
182	100.6	47.3	141.9
145	0.0	15.6	46.7



Abundance

Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

