

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082620\
 Data File : VN063142.D
 Acq On : 26 Aug 2020 18:59
 Operator : JC/MD
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Aug 27 02:12:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	152822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	252450	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	254638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	118246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96364	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
35) Dibromofluoromethane	7.55	113	83348	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	10.06	98	317238	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
62) 4-Bromofluorobenzene	12.38	95	118019	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
Target Compounds						
20) Methylene Chloride	4.50	84	3746	1.038	ug/l	95
95) Naphthalene	15.11	128	6107	6.791	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082620\
Data File : VN063142.D
Acq On : 26 Aug 2020 18:59
Operator : JC/MD
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Aug 27 02:12:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

