

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082621\
 Data File : VN068322.D
 Acq On : 26 Aug 2021 16:13
 Operator : JC/MD
 Sample : PB138716TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138716TB

Quant Time: Aug 27 05:53:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	426310	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	728350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	822039	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	345095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	241264	53.004	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.000%
35) Dibromofluoromethane	8.024	113	224825	49.166	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.340%
50) Toluene-d8	10.446	98	899226	51.609	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.220%
62) 4-Bromofluorobenzene	12.733	95	328398	56.617	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.240%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	38821	29.211	ug/l	# 89
20) Methylene Chloride	5.098	84	68071	17.116	ug/l	# 76
43) Isopropyl Acetate	8.568	43	41444	5.101	ug/l	# 90
95) Naphthalene	15.507	128	10686	4.080	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082621\
Data File : VN068322.D
Acq On : 26 Aug 2021 16:13
Operator : JC/MD
Sample : PB138716TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138716TB

Quant Time: Aug 27 05:53:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

