

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082621\
 Data File : VN068324.D
 Acq On : 26 Aug 2021 17:03
 Operator : JC/MD
 Sample : M3534-08RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210824RE

Manual Integrations
 APPROVED

SAM
 8/27/2021 1:34:55 PM

Quant Time: Aug 27 05:53:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	457320	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	792645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	806804	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	344954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	258085	52.855	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.700%			
35) Dibromofluoromethane	8.027	113	237472	47.719	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.440%			
50) Toluene-d8	10.446	98	923735	48.715	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.440%			
62) 4-Bromofluorobenzene	12.733	95	320482	50.771	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.540%			
Target Compounds						
11) Tert butyl alcohol	5.390	59	25715m	42.230	ug/l	Qvalue
16) Acetone	4.304	43	33296	23.355	ug/l #	86
95) Naphthalene	15.509	128	7696	3.875	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082621\
Data File : VN068324.D
Acq On : 26 Aug 2021 17:03
Operator : JC/MD
Sample : M3534-08RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-VPB-RW7-EB-20210824RE

Manual Integrations
APPROVED

SAM
8/27/2021 1:34:55 PM

Quant Time: Aug 27 05:53:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

