


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082621\
 Data File : VN068315.D
 Acq On : 26 Aug 2021 10:41
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Title : SW846 8260
 Last Update : Tue Aug 17 13:09:49 2021



AutoFind: Scans 4006, 4007, 4008; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	13440	PASS
75	95	30	60	51.0	40525	PASS
95	95	100	100	100.0	79509	PASS
96	95	5	9	6.6	5244	PASS
173	174	0.00	2	0.6	442	PASS
174	95	50	100	94.8	75339	PASS
175	174	5	9	7.3	5515	PASS
176	174	95	101	96.3	72517	PASS
177	176	5	9	6.4	4620	PASS