

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082624\
 Data File : VN083493.D
 Acq On : 26 Aug 2024 19:47
 Operator : JC\MD
 Sample : P3716-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1053-FB082024

Quant Time: Aug 27 02:02:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	111449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	227351	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	92412	58.254	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.500%
35) Dibromofluoromethane	8.165	113	75559	53.245	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.500%
50) Toluene-d8	10.565	98	257046	48.559	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.120%
62) 4-Bromofluorobenzene	12.853	95	118093	57.224	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	114.440%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	7795	12.558	ug/l	# 88
68) m/p-Xylenes	12.071	106	10155	2.723	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082624\
Data File : VN083493.D
Acq On : 26 Aug 2024 19:47
Operator : JC\MD
Sample : P3716-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1053-FB082024

Quant Time: Aug 27 02:02:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

