

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082718\
 Data File : VN050899.D
 Acq On : 27 Aug 2018 15:22
 Operator : MD\SY
 Sample : J4652-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SONO-S

Quant Time: Aug 28 03:19:20 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	615025	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1014504	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	861507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	310364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	409461	55.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.60%	
35) Dibromofluoromethane	7.59	113	381823	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	10.09	98	1386616	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
62) 4-Bromofluorobenzene	12.40	95	384577	38.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.70%	
Target Compounds						
16) Acetone	3.82	43	30355	10.55	ug/l	98
18) Methyl Acetate	4.34	43	17222	2.46	ug/l #	72
20) Methylene Chloride	4.55	84	77378	9.27	ug/l	97
30) Chloroform	7.37	83	24300	1.78	ug/l	94
43) Isopropyl Acetate	8.17	43	272048	23.01	ug/l	92
95) Naphthalene	15.14	128	19138	6.78	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082718\
Data File : VN050899.D
Acq On : 27 Aug 2018 15:22
Operator : MD\SY
Sample : J4652-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SONO-S

Quant Time: Aug 28 03:19:20 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

