

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
 Data File : VN068335.D
 Acq On : 27 Aug 2021 10:56
 Operator : JC/MD
 Sample : VN0827MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0827MBL01

Quant Time: Aug 28 01:25:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	480377	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	813309	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	837110	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	367449	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	266177	51.895	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.800%
35) Dibromofluoromethane	8.024	113	245876	48.153	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.300%
50) Toluene-d8	10.446	98	955156	49.092	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.180%
62) 4-Bromofluorobenzene	12.733	95	334605	51.661	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
Data File : VN068335.D
Acq On : 27 Aug 2021 10:56
Operator : JC/MD
Sample : VN0827MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0827MBL01

Quant Time: Aug 28 01:25:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

