

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
 Data File : VN068339.D
 Acq On : 27 Aug 2021 13:00
 Operator : JC/MD
 Sample : PB138747TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138747TB

Quant Time: Aug 28 01:26:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	387669	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	657497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	749770	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	318413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	223729	54.051	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	108.100%	
35) Dibromofluoromethane	8.024	113	206938	50.131	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	100.260%	
50) Toluene-d8	10.443	98	832176	52.907	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	105.820%	
62) 4-Bromofluorobenzene	12.734	95	303860	58.032	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	116.060%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	32779	27.123	ug/l	# 84
20) Methylene Chloride	5.095	84	59406	16.336	ug/l	# 79
43) Isopropyl Acetate	8.569	43	35466	4.835	ug/l	# 92
95) Naphthalene	15.509	128	19397	4.792	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
Data File : VN068339.D
Acq On : 27 Aug 2021 13:00
Operator : JC/MD
Sample : PB138747TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138747TB

Quant Time: Aug 28 01:26:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

