

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
 Data File : VN068340.D
 Acq On : 27 Aug 2021 13:25
 Operator : JC/MD
 Sample : M3550-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Aug 28 01:26:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	446901	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	780449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	884279	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	376321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	262871	55.090	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.180%			
35) Dibromofluoromethane	8.027	113	238333	48.641	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.280%			
50) Toluene-d8	10.443	98	990743	53.065	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.140%			
62) 4-Bromofluorobenzene	12.736	95	356666	57.386	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 114.780%			
Target Compounds						
						Qvalue
16) Acetone	4.293	43	81941	58.816	ug/l	# 88
20) Methylene Chloride	5.098	84	22628	3.910	ug/l	# 78
25) 2-Butanone	7.348	43	17614	8.113	ug/l	# 72
43) Isopropyl Acetate	8.566	43	35754	4.107	ug/l	# 93
95) Naphthalene	15.509	128	13377	4.189	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
Data File : VN068340.D
Acq On : 27 Aug 2021 13:25
Operator : JC/MD
Sample : M3550-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Aug 28 01:26:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

