

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
 Data File : VN068344.D
 Acq On : 27 Aug 2021 15:06
 Operator : JC/MD
 Sample : PB138768ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138768ZHE#02

Quant Time: Aug 28 01:28:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	441169	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	767406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	865381	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	373870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	234184	49.716	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.440%
35) Dibromofluoromethane	8.027	113	214220	44.463	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	88.920%
50) Toluene-d8	10.443	98	869349	47.355	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.700%
62) 4-Bromofluorobenzene	12.734	95	322316	52.741	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.480%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	41393	30.097	ug/l	# 85
20) Methylene Chloride	5.098	84	22085	3.841	ug/l	85
43) Isopropyl Acetate	8.566	43	36340	4.245	ug/l	# 91
95) Naphthalene	15.509	128	12634	4.148	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082721\
Data File : VN068344.D
Acq On : 27 Aug 2021 15:06
Operator : JC/MD
Sample : PB138768ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138768ZHE#02

Quant Time: Aug 28 01:28:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

