

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083501.D
 Acq On : 27 Aug 2024 11:52
 Operator : JC\MD
 Sample : PB162992TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162992TB

Quant Time: Aug 28 01:42:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	135266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278010	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	137229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108260	56.228	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.460%
35) Dibromofluoromethane	8.165	113	83858	48.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.660%
50) Toluene-d8	10.565	98	311428	48.112	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.220%
62) 4-Bromofluorobenzene	12.847	95	140346	55.615	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.220%
Target Compounds						
16) Acetone	4.430	43	19130	25.392	ug/l	97
37) Ethyl Acetate	7.559	43	3736	1.271	ug/l #	89
43) Isopropyl Acetate	8.688	43	240377m	46.958	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083501.D
Acq On : 27 Aug 2024 11:52
Operator : JC\MD
Sample : PB162992TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162992TB

Quant Time: Aug 28 01:42:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

