

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083506.D
 Acq On : 27 Aug 2024 13:52
 Operator : JC\MD
 Sample : PB162992ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162992ZHE#04

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

08/28/2024
 Supervised By :Mahesh
 Dadoda

Quant Time: Aug 28 01:45:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	D
-----							08
Internal Standards							
1) Pentafluorobenzene	8.224	168	116079	50.000	ug/l	0.00	
34) 1,4-Difluorobenzene	9.100	114	245027	50.000	ug/l	0.00	
63) Chlorobenzene-d5	11.865	117	279076	50.000	ug/l	0.00	
72) 1,4-Dichlorobenzene-d4	13.794	152	137737	50.000	ug/l	0.00	
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	8.577	65	111843	67.691	ug/l	0.00	
Spiked Amount	50.000	Range 74 - 125	Recovery	=	135.380%	#	
35) Dibromofluoromethane	8.165	113	88958	58.165	ug/l	0.00	
Spiked Amount	50.000	Range 75 - 124	Recovery	=	116.340%		
50) Toluene-d8	10.565	98	331346	58.079	ug/l	0.00	
Spiked Amount	50.000	Range 86 - 113	Recovery	=	116.160%	#	
62) 4-Bromofluorobenzene	12.847	95	154317	69.382	ug/l	0.00	
Spiked Amount	50.000	Range 77 - 121	Recovery	=	138.760%	#	
Target Compounds							
						Qvalue	
16) Acetone	4.436	43	20703	32.022	ug/l	96	
18) Methyl Acetate	5.018	43	11075	5.749	ug/l	95	
43) Isopropyl Acetate	8.688	43	274680m	62.055	ug/l		

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083506.D
Acq On : 27 Aug 2024 13:52
Operator : JC\MD
Sample : PB162992ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162992ZHE#04

Manual Integrations
APPROVED

Reviewed By :John
Carlone

08/28/2024
Supervised By :Mahesh
Dadoda

08/28/2024

