

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083513.D
 Acq On : 27 Aug 2024 16:42
 Operator : JC\MD
 Sample : P3727-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-21

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/28/2024
 Supervised By : Mahesh Dadoda 08/28/2024

Quant Time: Aug 28 08:24:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285137	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	314284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	156125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113532	56.260	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.520%			
35) Dibromofluoromethane	8.165	113	90821	51.030	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.060%			
50) Toluene-d8	10.565	98	324727	48.912	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.820%			
62) 4-Bromofluorobenzene	12.847	95	159159	61.493	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 122.980%#			
Target Compounds					Qvalue	
16) Acetone	4.436	43	20569	26.049	ug/l	100
18) Methyl Acetate	5.030	43	8723	3.708	ug/l #	78
20) Methylene Chloride	5.265	84	7456	4.103	ug/l #	80
43) Isopropyl Acetate	8.688	43	230958m	43.785	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083513.D
Acq On : 27 Aug 2024 16:42
Operator : JC\MD
Sample : P3727-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-21

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/28/2024
Supervised By :Mahesh Dadoda 08/28/2024

Quant Time: Aug 28 08:24:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

