

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083515.D
 Acq On : 27 Aug 2024 17:30
 Operator : JC\MD
 Sample : P3735-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-13

Quant Time: Aug 28 08:25:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137301	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	310325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	152500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110249	56.413	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.820%		
35) Dibromofluoromethane	8.165	113	87220	48.625	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.240%		
50) Toluene-d8	10.565	98	318786	47.643	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 95.280%		
62) 4-Bromofluorobenzene	12.847	95	153346	58.786	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 117.580%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	52822	69.073	ug/l	99
18) Methyl Acetate	5.024	43	8536	3.746	ug/l	99
20) Methylene Chloride	5.265	84	8307	4.720	ug/l #	73
25) 2-Butanone	7.488	43	6888	5.866	ug/l #	89
37) Ethyl Acetate	7.559	43	3552	1.169	ug/l #	91
43) Isopropyl Acetate	8.688	43	243650	45.981	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083515.D
Acq On : 27 Aug 2024 17:30
Operator : JC\MD
Sample : P3735-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-13

Quant Time: Aug 28 08:25:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

