

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083517.D
 Acq On : 27 Aug 2024 18:19
 Operator : JC\MD
 Sample : P3739-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Aug 28 08:25:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	137180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	282503	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154568	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108707	55.673	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.340%	
35) Dibromofluoromethane	8.165	113	86166	48.866	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.740%	
50) Toluene-d8	10.565	98	315818	48.014	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	96.020%	
62) 4-Bromofluorobenzene	12.847	95	153228	59.754	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	119.500%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	47753	62.500	ug/l	99
18) Methyl Acetate	5.024	43	8073	3.546	ug/l	100
20) Methylene Chloride	5.259	84	8640	4.914	ug/l #	80
25) 2-Butanone	7.494	43	6055	5.161	ug/l #	88
37) Ethyl Acetate	7.565	43	3423	1.146	ug/l #	88
43) Isopropyl Acetate	8.688	43	221138	42.210	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083517.D
Acq On : 27 Aug 2024 18:19
Operator : JC\MD
Sample : P3739-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Aug 28 08:25:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

