

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
 Data File : VN083520.D
 Acq On : 27 Aug 2024 19:32
 Operator : JC\MD
 Sample : P3739-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3

Quant Time: Aug 28 08:26:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275781	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	150037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	106751	56.327	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.660%		
35) Dibromofluoromethane	8.165	113	83532	48.527	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.060%		
50) Toluene-d8	10.565	98	309091	48.137	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.280%		
62) 4-Bromofluorobenzene	12.847	95	148501	59.322	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 118.640%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.506	59	2771	7.036	ug/l #	71
16) Acetone	4.436	43	50578	68.203	ug/l	99
18) Methyl Acetate	5.024	43	9261	4.191	ug/l #	81
20) Methylene Chloride	5.277	84	9612	5.632	ug/l	85
25) 2-Butanone	7.488	43	6263	5.500	ug/l #	78
43) Isopropyl Acetate	8.688	43	256782	50.839	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082724\
Data File : VN083520.D
Acq On : 27 Aug 2024 19:32
Operator : JC\MD
Sample : P3739-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3

Quant Time: Aug 28 08:26:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

