

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082818\
 Data File : VN050913.D
 Acq On : 28 Aug 2018 14:50
 Operator : MD\SY
 Sample : VN0828WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0828WBL01

Quant Time: Aug 28 15:10:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Aug 28 14:02:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	117585	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	606142	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	480012	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	228237	30.42	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.40%
60) 4-Bromofluorobenzene	12.40	95	158938	22.27	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.23%
63) Toluene-d8	10.09	98	698813	31.22	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.07%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082818\
Data File : VN050913.D
Acq On : 28 Aug 2018 14:50
Operator : MD\SY
Sample : VN0828WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0828WBL01

Quant Time: Aug 28 15:10:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Aug 28 14:02:24 2018
Response via : Initial Calibration

