

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082818\
 Data File : VN050930.D
 Acq On : 28 Aug 2018 22:21
 Operator : MD\SY
 Sample : J4277-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-082718-A

Quant Time: Aug 29 07:25:19 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	595249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	979071	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	831493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	296158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	397394	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
35) Dibromofluoromethane	7.60	113	371522	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
50) Toluene-d8	10.09	98	1328539	45.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.24%	
62) 4-Bromofluorobenzene	12.40	95	360192	37.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.44%	
Target Compounds						
16) Acetone	3.83	43	40932	17.55	ug/l	Qvalue 96
18) Methyl Acetate	4.34	43	13911	2.06	ug/l #	89
20) Methylene Chloride	4.55	84	55111	6.82	ug/l	94
30) Chloroform	7.38	83	21768	1.65	ug/l	100
43) Isopropyl Acetate	8.17	43	258525	22.66	ug/l	95
95) Naphthalene	15.13	128	77286	11.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082818\
Data File : VN050930.D
Acq On : 28 Aug 2018 22:21
Operator : MD\SY
Sample : J4277-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-01-082718-A

Quant Time: Aug 29 07:25:19 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

