

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079063.D
 Acq On : 28 Aug 2023 13:32
 Operator : JC\MD
 Sample : 04092-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL-DSN-001

Quant Time: Aug 29 05:21:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 18 05:21:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	36260	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	194043	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	167937	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	87704	29.856	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.533%
60) 4-Bromofluorobenzene	12.853	95	61337	25.038	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.467%
63) Toluene-d8	10.570	98	233109	29.724	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.067%
Target Compounds						
					Qvalue	
15) Acetone	4.436	58	282578	778.409	ug/l	94
25) Chloroform	7.971	83	124410	29.064	ug/l	98
41) Bromodichloromethane	9.888	83	6531	2.033	ug/l	96
53) Dibromochloromethane	11.370	129	3017	1.277	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079063.D
Acq On : 28 Aug 2023 13:32
Operator : JC\MD
Sample : 04092-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL-DSN-001

Quant Time: Aug 29 05:21:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 18 05:21:14 2023
Response via : Initial Calibration

