

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079067.D
 Acq On : 28 Aug 2023 15:18
 Operator : JC\MD
 Sample : PB155094TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155094TB

Quant Time: Aug 29 04:05:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	212905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	386093	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	375873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	158021	47.698	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.400%
35) Dibromofluoromethane	8.171	113	132615	51.026	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	10.571	98	488820	50.654	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.300%
62) 4-Bromofluorobenzene	12.853	95	153693	48.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.460%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	26311	19.247	ug/l	88
20) Methylene Chloride	5.271	84	8381	2.877	ug/l #	85
43) Isopropyl Acetate	8.694	43	156735	20.273	ug/l #	78
51) 4-Methyl-2-Pentanone	10.447	43	38084	8.305	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079067.D
Acq On : 28 Aug 2023 15:18
Operator : JC\MD
Sample : PB155094TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155094TB

Quant Time: Aug 29 04:05:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

