

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079069.D
 Acq On : 28 Aug 2023 16:07
 Operator : JC\MD
 Sample : 04170-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SCAS-01-G

Quant Time: Aug 29 04:06:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	206338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	380696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	368244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	158420	49.340	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.680%		
35) Dibromofluoromethane	8.177	113	135486	52.870	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	105.740%		
50) Toluene-d8	10.570	98	470408	49.437	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.880%		
62) 4-Bromofluorobenzene	12.853	95	148122	47.625	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.240%		
Target Compounds						
16) Acetone	4.441	43	24362	18.389	ug/l	91
20) Methylene Chloride	5.288	84	10013	3.547	ug/l #	61
43) Isopropyl Acetate	8.700	43	104180	13.666	ug/l #	77
51) 4-Methyl-2-Pentanone	10.447	43	33195	7.342	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079069.D
Acq On : 28 Aug 2023 16:07
Operator : JC\MD
Sample : 04170-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SCAS-01-G

Quant Time: Aug 29 04:06:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

