

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079072.D
 Acq On : 28 Aug 2023 17:19
 Operator : JC\MD
 Sample : 04188-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-5

Quant Time: Aug 29 04:40:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	208237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	394794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	386609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	158907	49.040	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.080%
35) Dibromofluoromethane	8.171	113	137840	51.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.740%
50) Toluene-d8	10.571	98	487564	49.410	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.853	95	158178	49.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.080%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	23376	17.484	ug/l #	87
20) Methylene Chloride	5.277	84	9432	3.310	ug/l #	72
43) Isopropyl Acetate	8.694	43	152459	19.285	ug/l #	76
51) 4-Methyl-2-Pentanone	10.453	43	35143	7.495	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079072.D
Acq On : 28 Aug 2023 17:19
Operator : JC\MD
Sample : 04188-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-5

Quant Time: Aug 29 04:40:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

