

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079074.D
 Acq On : 28 Aug 2023 18:08
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 29 07:54:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	150146	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	451951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	413032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	79254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	201424	86.211	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	172.420%#
35) Dibromofluoromethane	8.171	113	96935	31.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	63.720%#
50) Toluene-d8	10.570	98	557221	49.328	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.853	95	160125	43.367	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.740%
Target Compounds						
					Qvalue	
28) Bromochloromethane	7.824	49	63178	32.121	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079074.D
Acq On : 28 Aug 2023 18:08
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 29 07:54:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

