

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
 Data File : VN079075.D
 Acq On : 28 Aug 2023 18:32
 Operator : JC\MD
 Sample : 04204-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230815008-01

Quant Time: Aug 29 05:22:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 18 05:21:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	34219	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	191461	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	154253	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	86349	31.148	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.833%
60) 4-Bromofluorobenzene	12.853	95	52336	23.259	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	77.533%
63) Toluene-d8	10.571	98	223010	30.959	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.200%
Target Compounds						
25) Chloroform	7.965	83	51331	12.707	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082823\
Data File : VN079075.D
Acq On : 28 Aug 2023 18:32
Operator : JC\MD
Sample : 04204-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230815008-01

Quant Time: Aug 29 05:22:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 18 05:21:14 2023
Response via : Initial Calibration

