



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082824\
Data File : VN083523.D
Acq On : 28 Aug 2024 08:07
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N082824W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Aug 28 12:23:16 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	11407	PASS
75	95	30	60	56.4	30712	PASS
95	95	100	100	100.0	54472	PASS
96	95	5	9	6.7	3662	PASS
173	174	0.00	2	0.3	119	PASS
174	95	50	100	70.1	38171	PASS
175	174	5	9	7.5	2845	PASS
176	174	95	101	95.2	36333	PASS
177	176	5	9	7.1	2568	PASS